

## CHIK IgM ELISA Kit

**REF**

AE0315

**IVD**

- 96-well ELISA kit for the qualitative detection of IgM anti- Chikungunya virus in human serum or plasma
- Store at 2°C- 8°C upon receipt

### INTENDED USE

The TRUSTwell CHIK IgM ELISA Kit is a solid phase enzyme linked immunoabsorbent assay for the qualitative detection of IgM anti- Chikungunya virus in human serum or plasma. It is intended for professional use only as an aid in the diagnosis of acute infection with Chikungunya viruses. Any reactive specimen with the TRUSTwell CHIK IgM ELISA Kit must be confirmed with alternative testing method(s) and clinical findings.

### INTRODUCTION

Chikungunya is a rare viral infection transmitted by the bite of an infected *Aedes aegypti* mosquito. It is characterized by a rash, fever, and severe joint pain (arthralgias) that usually lasts for three to seven days. The name is derived from the Makonde word meaning "that which bends up" in reference to the stooped posture developed as a result of the arthritic symptoms of the disease. It occurs during the rainy season in tropical areas of the world, primarily in Africa, South-East Asia, southern India and Pakistan<sup>1-2</sup>.

The symptoms are most often clinically indistinguishable from those observed in dengue fever. Indeed, dual infection of dengue and chikungunya has been reported in India<sup>3</sup>. Unlike dengue, hemorrhagic manifestations are relatively rare and most often the disease is a self limiting febrile illness. Therefore it is very important to clinically distinguish dengue from CHIK infection.

CHIK is diagnosed based on serological analysis and viral isolation in mice or tissue culture. An IgM immunoassay is the most practical lab test method<sup>4</sup>.

The TRUSTwell CHIK IgM ELISA Kit utilizes recombinant CHIK antigens for detection of IgM antibodies associated with Chikungunya viruses.

### TEST PRINCIPLE

TRUSTwell CHIK IgM ELISA Kit is a solid phase enzyme linked immunoabsorbent assay based on the principle of the indirect immunoassay technique for the detection of IgM anti-Chikungunya virus in human serum or plasma.

The TRUSTwell CHIK IgM ELISA Kit is composed of two key components:

- 1) Solid microwells pre-coated with recombinant CHIK antigen(s);
- 2) Liquid conjugates composed of monoclonal anti-human IgM reagent conjugated with horse radish peroxidase (HRP-anti-Human IgM conjugates).

During the assay, the test specimen is first incubated with the coated microwells. IgM anti- Chikungunya virus, if present in the specimen, binds to the antigen coated on the microwell surface.

In the second incubation with the HRP-anti Human IgM conjugates, the IgM anti-Chikungunya antibody absorbed on the surface of microwell reacts to the conjugates, forming a complexed conjugates.

Unbounded conjugates are then removed by washing. The presence of the complexed conjugates is shown by a blue color upon additional incubation with TMB substrate. The reaction is stopped with Stop Solution and absorbances are read using a spectrophotometer at 450 /620-690 nm.

### MATERIALS AND REAGENTS

#### Materials and reagents provided with the kit

Item	Description	Quantity	Catalog
1	Microwells coated with recombinant CHIK antigens	8 wells x 12 strips	AE0315W
2	CHIK IgM positive control	0.5 mL	AE0315P
3	CHIK IgM negative control	0.5 mL	AE0315N
4	Sample diluent	12 mL	AE0315SD
5	HRP-anti-human IgM conjugate	12 mL	AE0315H
6	Wash buffer (30 x concentrate)	20 mL	AWE3000
7	TMB substrate A	6 mL	ATME2000A
8	TMB substrate B	6 mL	ATME2000B
9	Stop solution	12 mL	ASE1000
10	ELISA Working Sheet	2 Nos.	AE0001ES
11	Product insert	1 No.	PI-AE0315

#### Materials and reagents required but not provided in the kit

1. Pipette capable of delivering 10 µL, 50 µL and 100 µL volumes with a precision better than 1.5%.
2. Microplate reader with a bandwidth of 10 nm or less and an optical density range of 0-3 OD or greater at 450nm wavelength is acceptable
3. Absorbent paper for blotting the microplate wells.
4. Parafilm or other adhesive film sealant for sealing plate.
5. Timer.
6. Distilled or de-ionized water.

### STORAGE AND STABILITY

All reagents except the concentrated wash buffer are ready to use as supplied. Return all reagents requiring refrigeration immediately after use. Reseal the microwells after removing the desired number of wells. Ensure that the reagents are brought to room temperature before opening. All the reagents are stable through the expiration date printed on the label if not opened. Do not freeze the kit or expose the kit over 8°C.

### WARNING AND PRECAUTIONS

#### For in Vitro Diagnostic Use

1. This package insert must be read completely before performing the test. Failure to follow the insert gives inaccurate test results.
2. Do not use expired devices.
3. Bring all reagents to room temperature (18°C-28°C) before use.
4. Do not use the components in any other type of test kit as a substitute for the components in this kit.
5. Do not use hemolyzed blood specimen for testing.
6. Do not ingest the reagents. Avoid contact with eyes, skin and mucose. Wear protective clothing and disposable gloves while handling the kit reagents and clinical specimens. Wash hands thoroughly after performing the test.
7. Do not smoke, drink, or eat in areas where specimens or kit reagents are being handled.

8. Users of this test should follow the US CDC Universal Precautions for prevention of transmission of HIV, HBV and other blood-borne pathogens.
9. Dispose of all specimens and materials used to perform the test as biohazardous waste.
10. In the beginning of each incubation and after adding Stopping Solution, gently rocking the microwells to ensure thorough mixing. Avoid the formation of air bubbles as which results in inaccurate absorbance values. Avoid splash liquid while rocking or shaking the wells
11. Don't allow the microplate to dry between the end of the washing operation and the reagent distribution.
12. The enzyme reaction is very sensitive to metal ions. Thus, do not allow any metal element to come into contact with the conjugate or substrate solution.
13. The substrate solution must be colorless. The appearance of color indicates that the reagent cannot be used and must be replaced. The Substrate B must be stored in the dark.
14. Use a new distribution tip for each specimen. Never use the specimen container to distribute conjugate and substrate.
15. The wash procedure is critical. Wells must be aspirated completely before adding the Washing Solution or liquid reagents. Insufficient washing will result in poor precision and falsely elevated absorbance.
16. Avoid strong light during color development.

### SPECIMEN COLLECTION AND PREPARATION

- Serum or plasma should be prepared from a whole blood specimen obtained by acceptable venipuncture technique.
- This kit is designed for use with serum or plasma specimen without additives only.
- If a specimen is not tested immediately, refrigerated at 2°C-8°C. If storage period greater than three days are anticipated, the specimen should be frozen (-20°C). Avoid repeated freezing-thawing of specimens. If a specimen is to be shipped, pack in compliance with federal regulation covering the transportation of etiologic agents.
- Specimens containing precipitants may give inconsistent test results. Clarify such specimens by centrifugation prior to assaying.
- Do not use serum specimens demonstrating gross lipemia, gross hemolysis or turbidity. Do not use specimens containing sodium azide.

### PREPARATION OF THE REAGENTS

1. Bring all reagents, controls to room temperature (18°C-28°C).
2. Dilute concentrated Wash Buffer 30 fold with water as following:

Plate	DI water	30 X wash buffer	Final volume
Full plate	580 mL	20 mL	600 mL
Half plate	290 mL	10 mL	300 mL
A quarter plate	145 mL	5 mL	150 mL

Warm up the concentrated Washing Buffer at 37°C to dissolve the precipitant if it appears.

3. Mix each reagent before adding to the test wells.
4. Determine the number of microwells needed and mark on the ELISA Working Sheet with the appropriate information. Positive and Negative Controls require to be run in duplicate to ensure accuracy.

## ASSAY PROCEDURE

- Remove the desired number of strips and secure them in the microwell frame. Reseal un-used strips.
- Add specimens according to the designation on the ELISA Working Sheet
  - Blank well:** Leave the blank well alone. Don't add any reagents.
  - Control wells:** Add 100 µL of Positive, Negative Control into the designated control wells, respectively.
  - Test wells:** Add 100 µL of Sample Diluent to all the test wells, then transfer 10 µL of each test specimen to each test well, respectively.

To ensure better precision, use pipette to handle solution.

- Gently rock the plate wells for twenty second, then cover the plate with sealant.
- Incubate the wells at 37°C for 45 minutes.
- Carefully remove the incubation mixture by emptying the solution into a waste container. Fill each well with diluted wash buffer and shake gently for 20-30 second. Discard the wash solution completely and tapping the plate on absorbent paper. Repeat above procedure 4 more times.
- Add 100 µL of HRP- anti-human IgM conjugates into each well except the blank well, cover the plate, and incubate at 37°C for 30 minutes.
- Wash the plate 5 times as step 5 described.
- Add 50 µL of TMB substrate A and 50 µL of TMB substrate B into each well including the blank well.
- Incubate at 37°C in dark for 20 minutes.
- Stop the reaction by adding 100 µL of stop buffer to each well. Gently mix for 30 seconds. **It is important to make sure that all the blue color changes to yellow color completely.**
- Set the microplate reader wavelength at 450nm and measure the absorbance (OD) of each well against the blank well within 15 minutes after adding Stop Solution. A filter of 620 -690nm can be used as a reference wavelength to optimize the assay result.

## INTERPRETATION OF RESULTS

### A. Set up the cut-off value

The cut-off value = 0.30 + N  
N: Mean OD of the negative control.

### B. Calculation of specimen OD ratio

Calculate an OD ratio for each specimen by dividing its OD value by the cut-off Value as follows:

$$\text{Specimen OD ratio} = \frac{\text{Specimen OD}}{\text{Cut-off Value}}$$

### C. Assay Validation

The mean OD value of the positive controls should be  $\geq 0.50$ .  
The mean OD value of the negative controls should be  $\leq 0.10$ .

Check the procedure and repeat assay if above conditions are not met.

### D. Interpretation of the results

#### Specimen OD ratio

Negative	< 1.00
Positive	$\geq 1.00$

- The negative result indicates that there is no detectable Chikungunya IgM antibody in the specimen.
- Results just below the cut-off value (Lower than 10% of the cut-off value) should be interpreted with caution (it is advisable to retest in duplicate the corresponding specimens when it is applicable). Specimens, which are repeatedly borderline should be tested using an alternative method such as IFA test system.
- Specimens with cut-off  $\geq 1.00$  are initially considered to be positive by the TRUSTwell CHIK IgM ELISA Kit. They should be retested in duplicate before final interpretation.

If after re-testing of a specimen, the absorbance value of the 2 duplicates are less than the cut-off value, the initial result is non repeatable and the specimen is considered to be negative with the TRUSTwell CHIK IgM ELISA Kit.

Non repeatable reactions are often caused by:

- Inadequate microwell washing,
- Contamination of negative specimens by serum or plasma with a high antibody titer,
- Contamination of the substrate solution by oxidizing agents (bleach, metal ions, etc.)
- Contamination of the stopping solution

If after retesting the absorbance of one of the duplicates is equal or greater than the cut-off value, the initial result is repeatable and the specimen is considered to be positive with the TRUSTwell CHIK IgM ELISA Kit, subject to the limitation of the procedure, described below.

## PERFORMANCE CHARACTERISTICS

### Clinical Performance

A total of 336 specimens from susceptible subjects were tested by TRUSTwell CHIK IgM ELISA and by a commercial reference ELISA kits. Comparison for all subjects is showed in the following table:

Reference EIA	TRUSTwell CHIK IgM ELISA Kit		
	Positive	Negative	Total
Positive	21	2	23
Negative	10	303	313
Total	31	305	336

Relative Sensitivity: 91.3 %, (95 % CI: 73.2 % -97.6 %)  
Relative Specificity: 96.8 %, (95 % CI: 94.2 % -98.3 %)  
Overall Agreement: 96.4%, (95 % CI: 93.9 % -98.0 %)

## LIMITATION OF THE TEST

- The Assay Procedure and the Test Result Interpretation must be followed closely when testing the presence of antibodies to CHIK virus in serum or plasma from individual subjects. Failure to follow the procedure may give inaccurate results.
- The TRUSTwell CHIK IgM ELISA Kit is limited to the qualitative detection of IgM anti- Chikungunya in human serum or plasma. It is not a quantitative test. The intensity of color does not have linear correlation with the antibody titer in the specimen.
- A negative result for an individual subject indicates absence of detectable IgM anti- Chikungunya. However, a negative test result does not preclude the possibility of exposure to or infection with Chikungunya virus.
- A negative result can occur if the quantity of IgM anti-Chikungunya present in the specimen is below the detection limits of the assay, or the antibodies that are detected are not present during the stage of disease in which a sample is collected.
- Serological cross reactivity with other alphaviruses is reported, therefore, it is possible that patients infected with these viruses may show some level of the reactivity with this test.
- Some specimens containing unusually high titer of heterophile antibodies or rheumatoid factor may affect expected results.
- The results obtained with this test should only be interpreted in conjunction with other diagnostic procedures and clinical findings.

## REFERENCES

- Shah KV, Gibbs CJ Jr, Banerjee G. Virological investigation of the epidemic of haemorrhagic fever in Calcutta: isolation of three strains of Chikungunya virus. Indian J Med Res 1964; 52 :676-83.
- Powers AM, Brault AC, Tesh RB, Weaver SC. Re-emergence of Chikungunya and O'nyong-nyong viruses: evidence for distinct geographical lineages and distant evolutionary relationships. J Gen Virol 2000;81:471-9
- Myers RM and Carey DE. Concurrent isolation from patient of two arboviruses, Chikungunya and dengue type 2. Science 1967;157:1307-8.
- Thein S, La Linn M, Aaskov J, Aung MM, Aye M, Zaw A, Myint A. Development of a simple indirect enzyme-linked immunosorbent assay for the detection of immunoglobulin M antibody in serum from patients Following an outbreak of chikungunya virus infection in Yangon, Myanmar. Trans R Soc Trop Med Hyg. 1992 86:438-42.
- Yamamoto K, Hashimoto K, Ogata T Structural proteins of Chikungunya virus. Simizu B, J Virol. 1984 Jul;51(1):254-8

## Index of Symbols

	Consult instructions for use		Catalogue number		Use-by date
	For <i>in vitro</i> diagnostic use only		Batch code		Tests per kit
	Temperature limit 2-8°C		Do not re-use		Keep dry
	Manufacturer		Date of manufacture		If device is non-sterile
	Do not use if package is damaged		Keep away from sunlight		Warnings / Precautions

**ATHENESE-Dx**  
Athenese-Dx Pvt. Ltd.  
Module No. 407 & 408, 4<sup>th</sup> Floor,  
TICEL Bio Park II, No. 5, CSIR Road,  
Taramani, Chennai-600113, India  
Tel: +91-44-22541131  
E-mail: info@athensedx.com  
Website: www.athensedx.com

PI-AE0315 Rev. D  
Effective date: 01.09.2022  
English version