

Free Triiodothyronine (fT3) ELISA

IVD REF AE1016

- 96-well ELISA kit for the quantitative determination of Free Triiodothyronine (fT3) concentration in human serum
- Store at 2-8°C upon receipt

INTENDED USE

The Free Triiodothyronine (fT3) ELISA is a competitive solid-phase enzyme-linked immunosorbent assay for the quantitative determination of fT3 concentration in human serum. It is intended to be used by professionals as an aid in the diagnosis of thyroid dysfunction.

INTRODUCTION

Thyroid disease is very common on a global level and studies have shown that the prevalence of undiagnosed thyroid disorder is high¹⁻³. Thus, quantification of markers that reflect thyroid function is very important. Thyroid function and metabolism are each individual by the human hypothalamic-pituitary-thyroid (HPT) axis, which is made up of the hypothalamus, the pituitary gland, and the thyroid. Thyroid stimulating hormone (TSH) controls the production of the thyroid hormones thyroxine (T4) and 3,5,3' triiodothyronine (T3), which play an important role in the regulation of metabolism⁴⁻⁶.

The majority of T3 and T4 that circulates in the blood is protein-bound and attaches to plasma proteins such as Thyroxine-Binding Globulin (TBG). The small percentage that is not protein-bound is considered "free" (fT4/fT3), and can readily enter target tissue. Importantly, the binding protein concentration of each individual can be affected by factors such as pregnancy, hormonal therapy, and certain medications, which will, therefore, affect T4/T3 concentrations⁶⁻⁸. Due to the general stability of the "free" thyroid hormone concentrations, fT4/fT3 test results may be considered more useful in many cases^{3,6,7,8}.

Testing of thyroid function commonly incorporates TSH testing along with a combination of T3, T4, fT3, and fT4 assays^{3,7,8}. Generally, and depending on the status of the individual's TSH concentration, elevated fT3 results may indicate hyperthyroidism, and low fT3 results may indicate hypothyroidism. Free T4 tests are also commonly used to monitor pregnancy and certain treatments for thyroid disorder, pituitary disease, and thyroid cancer^{8,9}. Combining fT3 with other thyroid hormone results and any other necessary test results will help provide enough information for the healthcare provider to design a proper plan of care.

TEST PRINCIPLE

The Free Triiodothyronine (fT3) ELISA is a competitive solid-phase enzyme-linked immunosorbent assay for the quantitative measurement of fT3 concentration in human serum.

The Free Triiodothyronine (fT3) ELISA is comprised of three key components:

- 1) Solid microwells pre-coated with anti-fT3 antibody,
- 2) fT3 Calibrators
- 3) fT3 Enzyme Reagent comprised of horseradish peroxidase conjugated to T3 (T3- HRP)

During the assay, the calibrator, control, or test specimen as well as the T3-HRP are added to the T3-antibody coated microwell. During the 60-minute incubation, the fT3 in the test specimen and in the T3-HRP compete for binding to the antibody coated on the microwell surface.

Unbound material is then removed by washing. In the next step, the TMB Substrate is added and the presence of the HRP bound to microwell surface is shown by the development of a blue color. The reaction is then terminated with the Stop Solution and the optical density (OD) is determined using a spectrophotometer at 450/610-650 nm. The color intensity reflects the amount of T3-HRP bound to the microwell surface, and is inversely proportional to the amount of fT3 in the control or test specimen.

A standard curve can then be developed by plotting the fT3 Calibrator concentrations on the x-axis against the relative OD values on the y-axis. The fT3 concentration of each assay control or test specimen can then be interpolated from the standard curve.

MATERIALS AND REAGENTS

Materials and reagents provided:

Item	Description	Quantity	Catalog
1	Anti-fT3 Ab Coated Microwells	8 wells x 12 strips	AE1016W
2	fT3 Calibrators* C1 (0.00 pg/mL)	1 mL	AE1016C1
3	C2 (2.20 pg/mL)	1 mL	AE1016C2
4	C3 (4.50 pg/mL)	1 mL	AE1016C3
5	C4 (7.20 pg/mL)	1 mL	AE1016C4
6	C5 (11.0 pg/mL)	1 mL	AE1016C5
7	C6 (20.00 pg/mL)	1 mL	AE1016C6
8	fT3 Assay Control*	1 mL	AE1016AC
9	fT3 Enzyme Reagent	12 mL	AE1016H
10	Wash Buffer Concentrate (40X)	20 mL	AWE3002
11	TMB Substrate	12 mL	ATME2000
12	Stop Solution	7 mL	ASE1000
13	Product Insert	1	PI-AE1016
14	ELISA Working Sheet	2	AE001ES
15	Microplate Sealers	3	
16	Desiccant	4	

*The Calibrators and Assay Controls provided are lot-specific. Please see vial labels for exact concentrations or ranges, respectively.

Materials and reagents required but not provided in the kit

1. Pipette capable of delivering 50 µL, 100 µL, and 1 mL
2. Microplate reader, washer and incubator
3. Vortex mixer or equivalent
4. Absorbent paper for blotting the microwells
5. Graph paper
6. Timer
7. Distilled or de-ionized water

STORAGE AND STABILITY

All reagents except the concentrated wash buffer are ready to use as supplied. Store all components at 2-8°C. Do not freeze. Avoid strong light. Ensure that the reagents are brought to room temperature before opening. After removing the desired number of wells, place unused wells in the resealable plastic bag provided with desiccant and return to 2-8°C. All reagents are stable through the expiration date printed on the label if not opened. Once opened, the kit is stable for 8 weeks at 2-8°C, or until the labeled expiration date, whichever is earlier.

SPECIMEN COLLECTION AND PREPARATION

- Serum specimens should be prepared from whole blood obtained by acceptable venipuncture technique.
- If not tested immediately, the specimens can be stored at 2-8°C for up to 7 days. The specimens should be frozen at -20°C for longer storage. Avoid multiple freeze-thaw cycles. If a specimen is to be shipped, pack in compliance with federal regulations covering the transportation of etiologic agents.
- Specimens containing precipitants may give inconsistent test results. Clarify such specimens by centrifugation before testing.
- Do not use specimens demonstrating gross lipemia, gross hemolysis or turbidity.
- Specimens containing sodium azide may interfere with test results.

PREPARATION OF THE REAGENTS

1. Bring all reagents to room temperature (22-28°C).
2. **Preparation of working Wash Buffer:**
If precipitants are visible, warm up the Wash Buffer (40X concentrate) at 37°C. Dilute concentrated Wash Buffer 40 fold with water as follows:

Plate	Dl water	40X wash buffer	Final volume
Full plate	195 mL	5 mL	200 mL
Half plate	97.5 mL	2.5 mL	100 mL
Quarter plate	48.75 mL	1.25 mL	50 mL

3. Mix each reagent before adding to the test wells.
4. Determine the number of strips needed and mark on the ELISA working sheet with the appropriate information. **Calibrators should be run in duplicate to ensure accuracy.**











ASSAY PROCEDURE

1. Calculate the desired number of microwells. Remove the remaining microwells

and place them with desiccant into the resealable foil bag, seal and store at 2-8°C for later use.

2. Add 50 µL of fT3 calibrators, fT3 Assay Controls, and test specimens into the assigned wells, respectively.
3. Add 100 µL of fT3 Enzyme Reagent into all wells.
4. Cover the plate with a microplate sealer, and then shake gently for 10 seconds.
5. Incubate the microplate at room temperature (22-28°C) for 60 minutes.
6. Wash Step (Can be performed manually or with automated washing):
Manual washing: Carefully remove the incubation mixture by disposing the solution into a waste container. Fill each well with 350 µL of working wash buffer and mix gently for 20-30 seconds. Discard the wash solution completely. Repeat 2 more times. After completing the last wash step, tap the plate on absorbent paper to remove residual liquid.
Automated washing: Automatic plate washer must be calibrated to ensure efficient washing. Fill each well with 350 µL of diluted wash buffer. Aspirate all wells completely. Repeat 2 more times.
7. Add 100 µL of TMB Substrate into each well.
8. Cover the microplate and incubate at room temperature (22-28°C) for 15 minutes.
9. Stop the reaction by adding 50 µL of Stop Solution to each well. Gently mix for 10 seconds. Add the Stop Solution in the same sequence as substrate addition. **It is important to make sure that all the blue color changes completely to a yellow color.**
10. Set the microplate reader wavelength at 450 nm. Measure the OD value of each well within 30 minutes after adding Stop Solution. A filter of 610-650 nm can be used as a reference wavelength to optimize the assay result.

Flow chart of assay procedure

1.	Secure strips in microwell frame		Number of strips
2.	Add fT3 Calibrators, fT3 Assay Controls, or specimens		50 µL
3.	Add fT3 Enzyme Reagent		100 µL
4.	Shake		10 seconds
5.	Incubate		22-28°C, 60 minutes
6.	Wash: manual or automatic		3 times
7.	Add TMB Substrate		100 µL
8.	Incubate in dark		22-28°C, 15 minutes
9.	Add Stop Solution Gently mix		50 µL 10 seconds
10.	Read result		450/610-650 nm within 30 minutes

CALCULATION OF RESULTS

1. Calculate the mean OD value (A450/610-650 nm) for each set of calibrators.
2. Construct a standard curve by plotting the mean OD value obtained for each calibrator against its concentration on graph paper with OD values on the vertical y-axis, and concentrations on the horizontal x-axis.
3. Interpolate concentrations for the assay controls and test specimens by plotting their mean OD values on the curve.
Alternatively, if software is used, calculate the concentration of fT3 following


the software menu.

- If the OD value of a specimen is greater than that of the highest calibrator, it is recommended to dilute the specimen with C1 in 1:1 or 1:3 ratio and test again. Any values obtained for a diluted sample must be further converted by applying the appropriate dilution factor in the calculation.

INTERPRETATION

- Results of a typical standard curve are shown below:

Cal ID	Conc. (pg/mL)	OD
C1	0.0	2.574
C2	2.20	2.282
C3	4.50	1.847
C4	7.20	1.096
C5	11.00	0.840
C6	20.00	0.365
Patient ID	Conc. (pg/mL)	OD
1	6.00	1.470



The above data and figure are for example purposes and should not be used to calculate your result.

QUALITY CONTROL

Good laboratory practice requires that quality control specimens (controls) be run with each calibration curve to check assay performance. Any material used should be assayed repeatedly to establish mean values and acceptable ranges to assure proper performance.

The following criteria should be met to consider assay results to be valid:

- The mean OD of the '0' calibrator should be ≥ 1.3 .
- CV's for duplicates of calibrators should be $\leq 15\%$.
- FT3 Assay Control concentrations should be within the specified range on the vial labels.

NORMAL REFERENCE

It is recommended that each laboratory establish its own normal ranges based on a representative sampling of the local patient population and its own assay technique. The following values for the Free Triiodothyronine (FT3) ELISA can be used as initial guideline ranges only:

Normal Values for FT3 (pg/mL)

Sample Number (n)	106
Average	2.53
Standard Deviation (SD)	0.75
Range ($\pm 2SD$)	1.03-4.03

PERFORMANCE CHARACTERISTICS

1. Analytical Sensitivity

Twenty (20) replicates of the '0' calibrator were run in a single assay on each of three lots of the Free Triiodothyronine (FT3) ELISA, and the mean and standard deviation (SD) of the OD values were calculated. The analytical sensitivity of the test was interpolated from the standard curve. The analytical sensitivity of the Free Triiodothyronine (FT3) ELISA was determined to be 0.982 pg/mL at 2SD.

2. Specificity

Specificity of the Free Triiodothyronine (FT3) ELISA was evaluated by spiking potential interfering substances into a human serum matrix. Specificity was calculated by deriving a ratio between the concentration of each tested metabolite and the concentration of T3 needed to displace the same amount of conjugate. There was no cross-reactivity observed at the concentrations listed below:

d-Thyroxine	10 µg/mL
d-Triiodothyronine	100 µg/mL
Iodotyrosine	100 µg/mL
Diiodotyrosine	100 µg/mL
TBG	50 µg/mL
Human Albumin	50 mg/mL
Phenylbutazone	25 µg/mL
Phenytoin	50 µg/mL
ASA	500 µg/mL
Acetaminophen	500 µg/mL

3. Accuracy

The accuracy of the Free Triiodothyronine (FT3) ELISA was determined by external comparison of 65 specimens with varying concentrations of FT3 with a reference commercial ELISA. The results are listed in the following table:

Method	Mean (pg/mL)	Least Square Regression	Corr. Coef. (r)
(x)	3.27	$Y = 0.8393(x) + 0.612$	0.91
Reference (y)	3.17		

4. Precision

- Intra-assay precision:** Twenty replicates of each of three pooled human serum controls (low, medium, and high concentrations) were tested and the mean, SD, and coefficient of variation (CV%) were determined. The results are shown in the following table:

Sample	N	Mean (pg/mL)	SD	CV
Low	20	6.50	0.206	3.20
Medium	20	8.45	0.372	4.36
High	20	12.11	0.436	3.73

- Inter-Assay Precision:** Three human serum pooled controls (low, middle, and high concentrations) were assayed in 20 separate runs over 3 lots. The mean, SD, and CV% were determined. The results are shown in the following table:

Sample	N	Mean (pg/mL)	SD	CV
Low	20	3.10	0.40	13.0
Medium	20	5.50	0.57	10.3
High	20	7.70	0.95	12.3

WARNING AND PRECAUTIONS

For in Vitro Diagnostic Use

- This package insert must be read completely before performing the test. Failure to follow the instructions in the insert will lead to inaccurate test results.
- Do not use expired test kits.
- Bring all reagents to room temperature (22-28°C) before use.
- Do not use a component from any other test kit as a substitute for the components in this kit.
- Do not use serum derived from hemolyzed blood specimens for testing.
- Do not ingest the reagents. Avoid contact with eyes, skin and mouth. Wear protective clothing and disposable gloves while handling the kit reagents and clinical specimens. Wash hands thoroughly after performing the test.
- Do not smoke, drink, or eat in areas where specimens or kit reagents are being handled.
- Users of this test should follow the US CDC Universal Precautions for prevention of transmission of HIV, HBV and other blood-borne pathogens.
- All specimens and materials used to perform this test must be disposed of as bio-hazardous waste.
- Prior to the first incubation, and after addition of the Stop Solution, gently shake the microwells to ensure thorough mixing. Avoid splashing liquid while shaking. Do not allow formation of air bubbles in the microwell.
- Do not allow the microwells to dry between the end of the washing operation and the reagent distribution.
- The enzyme substrate reaction is very sensitive to metal ions. Thus, do not allow any metal elements to come into contact with the conjugate or TMB Substrate.
- The enzyme-substrate reaction is temperature dependent. Ensure that the room temperature is between 22-28°C during the TMB incubation.
- The TMB substrate must be colorless. The appearance of color indicates that the reagent cannot be used and must be replaced. The TMB Substrate must be stored in the dark.
- Use a new dispensing tip for each specimen. Never use the specimen container to distribute conjugate and TMB Substrate.
- The wash procedure is critical. Wells must be aspirated completely before adding the Wash buffer or liquid reagents. Automatic washers must be validated with the test kit prior to use. Insufficient washing will result in poor precision and falsely elevated OD values.**
- Microplate reader must be calibrated per manufacturer's instruction to ensure accurate OD readings. A non-calibrated reader may lead to invalid test results.**
- Avoid exposure to strong light during color development.

LIMITATIONS OF TEST

- The Assay Procedure and the Interpretation of Results must be followed closely

when testing for FT3 concentration in human specimens. Failure to follow the procedure may lead to inaccurate results.

- The test is limited to the detection of FT3 concentration in human serum.
- A clinical diagnosis should not be based on the results of a single test and should only be made by the physician after all clinical and laboratory findings have been evaluated.
- Normal thyroid hormone levels do not exclude thyroid disease.
- If the OD value of a specimen is greater than that of the highest calibrator, it is recommended to dilute the specimen with Calibrator C1 and test again.
- Any interpretation or use of this test result must also integrate other clinical findings as well as on the professional judgment of health care providers.

REFERENCES

- Soldin OP et al. Trimester-specific changes in maternal thyroid hormone, thyrotropin, and thyroglobulin concentrations during gestation: trends and associations across trimesters in iodine sufficiency. *Thyroid* 2014; 14:1084-1090.
- Vanderpump MPJ. The epidemiology of thyroid disease. *Br Med Bull* 2011; 99:39-51.
- Dufour DR. Laboratory tests of thyroid function: uses and limitations. *Endocr Metab Clin North Am* 2007; 36:155-169.
- Ezrin C, The Thyroid, S. C. Werner and S. H. Ingbar (eds.), Harper and Row, Hagerstown, MD. 1978; 9: 174-178.
- Boelaert K and Franklyn JA. Thyroid hormone in health and disease. *J. Endocrinol* 2005; 187:1-15.
- Spencer CA. Assay of Thyroid Hormones and Related Substances. In: De Groot LJ, Chrousos G, Dungan K, et al. *Endotext*. South Dartmouth (MA): MDText.com, Inc. 2000.
- Werner SC et al. Werner and Ingbar's the thyroid: a fundamental and clinical text. In: Werner and Ingbar's The Thyroid: a Fundamental and Clinical Text. Philadelphia: Lippincott-Raven. 1996.
- Sterling L, Diagnosis and treatment of Thyroid Disease, Cleveland, CRC Press. 1975; 9, 51.
- Wiens FH. Clinical Laboratory Tests: Which, Why, and What Do The Results Mean? *Lab Med* 2009; 40(2):105-113.

Index of Symbols

	Consult instructions for use		Catalogue number		Tests per kit
	For in vitro diagnostic use only		Batch code		Use-by date
	Temperature limit 2-8 °C		Do not re-use		Keep dry
	Manufacturer		Date of manufacture		Keep away from sunlight
	If device is non-sterile		Warnings / Precautions		

PI-AE1016 Rev. A
Effective date: 20.12.2021
English version

Athenese-Dx Pvt. Ltd.
Module No. 407 & 408, 4th Floor,
TICEL Bio Park II, No. 5, CSIR Road,
Taramani, Chennai-600113, India
Tel: +91-44-22541131
E-mail: info@athensedx.com
Website: www.athensedx.com